

Minutes of the 15th Inter-Ministerial Committee (IMC) to consider applications for recognition as Pre-shipment Inspection Agency (PSIA)

The 15th Meeting of the IMC was held on 26.06.2020 on video-conferencing via webex under the Chairmanship of Shri Hardeep Singh, Addl. DGFT to consider (i) the cases deferred in the 14th Meeting of PSIA held on 21.06.2019 (ii) fresh applications received for recognition of PSIA and (iii) extension of period of validity as recognized PSIA of those PSIA's whose term is ending by June, 2020.

2. The following members attended the meeting :

- I. Shri Rajbir Sharma, Joint DGFT
- II. Dr. Avinash Sonawane, Head, RSD, AERB
- III. Shri Ajeet Singh, SO (D), AERB
- IV. Shri Nand Lal, Consultant, DPIIT
- V. Shri Nirmal Kumar, Dy. DGFT
- VI. Shri Abdul Sadiq Khan, FTDO
- VII. Mrs. Pratibha Kumari, FTDO

3. A total number of 23 applications were received. Out of 23 applications, 22 applications were for empanelment of PSIA and extension of period of validity as PSIA and 01 was for addition of instruments and countries. The committee noted that there are deficiencies in all the applications. Accordingly, the Committee decided that the deficiencies in the applications be conveyed to the applicants so that they can rectify the same and re-submit within a period of 30 days i.e. by 17.08.2020. After examination of the fresh documents submitted, applications will be reconsidered by the Committee. The deficiency in the applications is indicated against the name of each applicant in the Appendix.

4. The Committee ended with the vote of thanks to the chair.

S. No.	Name of the Applicant	Make	Model	SI. No.	Date of calibration	Expiration Date of Calibration Certificate	Supporting documents of calibration	Date of Purchase	Copy of invoice should be attached	photograph attached	AERB Comments	Remark of IMC meetings
1	3	4	5	6	7		8	9	10	11	14	13
1	SMV International Incorporation	Mineralab LLC	Rad100X	X01737	11.01.2017	10.01.2018	Yes	16.01.2017	Yes	Yes	Calibration Certificate Expired	The firm is to submit revalidated calibration certificate of all instruments and revalidated certificate of membership of Metal Recycling Association of India
		Mineralab LLC	Rad100X	X01738	11.01.2017	10.01.2018	Yes	16.01.2017	Yes	Yes	Calibration Certificate Expired	
		Mineralab LLC	Rad100X	X01739	11.01.2017	10.01.2018	Yes	16.01.2017	Yes	Yes	Calibration Certificate Expired	
		Mineralab LLC	Rad100X	X01740	11.01.2017	10.01.2018	Yes	16.01.2017	Yes	Yes	Calibration Certificate Expired	
		Mineralab LLC	Rad100X	X01741	11.01.2017	10.01.2018	Yes	16.01.2017	Yes	Yes	Calibration Certificate Expired	
		Mineralab LLC	Rad100X	X01742	11.01.2017	10.01.2018	Yes	16.01.2017	Yes	Yes	Calibration Certificate Expired	
		Mineralab LLC	Rad100X	X01743	11.01.2017	10.01.2018	Yes	16.01.2017	Yes	Yes	Calibration Certificate Expired	
		Mineralab LLC	Rad100X	X01744	11.01.2017	10.01.2018	Yes	16.01.2017	Yes	Yes	Calibration Certificate Expired	
		Mineralab LLC	Rad100X	X01745	11.01.2017	10.01.2018	Yes	16.01.2017	Yes	Yes	Calibration Certificate Expired	
		Mineralab LLC	Rad100X	X01746	11.01.2017	10.01.2018	Yes	16.01.2017	Yes	Yes	Calibration Certificate Expired	
		Mineralab LLC	Rad100X	X01747	11.01.2017	10.01.2018	Yes	16.01.2017	Yes	Yes	Calibration Certificate Expired	
		Mineralab LLC	Rad100X	X01748	11.01.2017	10.01.2018	Yes	16.01.2017	Yes	Yes	Calibration Certificate Expired	
		RAD/COMM System	Mspec	200268	16.01.2017	15.01.2018	Yes	15.01.2017	Yes	Yes	Calibration Certificate Expired	The firm is to submit photograph of explosive detector
		Morphix Technologies	TraceX	510100-05	NR	NR	NR	12.01.2017	Yes	NA	Explosive Detector	

2	Gattini & Co.	RadComm Systems, 2931 Portland Drive, Oakville, Ontario,	Rhandy	100379	31.08.2017	-	Yes	11.08.2015	Yes	Yes	Calibration Certificate Expired	The firm is to submit revalidated calibration certificate clearly indicating its' validity of all instruments and revalidated certificate of membership of Metal Recycling Association of India,
		RadComm Systems, 2931 Portland Drive, Oakville, Ontario, Canada L6H 5S4	Rhandy	100512	13.08.2015	-	Yes	13.10.2015	Yes	Yes	Calibration Certificate Expired	
		RadComm Systems, 2931 Portland Drive, Oakville, Ontario, Canada L6H 5S4	Rhandy	100514	12.01.2016	-	Yes	13.01.2016	Yes	Yes	Calibration Certificate Expired	
		Quarta-Rad, Inc. 1201 Orange St, Ste 700, Wilmington, DE 19801	Radex RD1212	1.11601E+13	07.03.2017	-	NA	06.02.2017	Yes	Yes	Calibration Certificate Expired	
		Spectrometer	MSpec	200051	04.12.2018	04.12.2019	Yes	22.07.2013	Yes	Yes	Calibration Certificate Expired, but valid on date of application submission	
		Spectrometer	Mspec	200043	04.12.2018	04.12.2019	Yes	11.08.2015	Yes	Yes	Calibration Certificate Expired, but valid on date on application submission	The firm is to submit photograph of explosive detector
		Explosive Detection	TraceX	510100	NR	NR	Yes	31.08.2016	Yes	NA		

3	Skanda Engineers and Consultants	Radcomm, Canada	RHANDY	100655	19.09.2019	18.09.2020	Yes	21.09.2019	Yes	Yes		The firm is to submit revalidated certificate of membership of Metal Recycling Association of India
		Radcomm, Canada	Mspec	200341	19.09.2019	19.09.2020	Yes	21.09.2019	Yes	Yes		
		Explosives Detection Morphix Technologies	TraceX	16858 Part No 510100	NR	NR	NR	20.09.2019	Yes	Yes		
4		RADCOMM	Rhandy Portable	100397	06.05.2019	05.05.2020	Yes	03.02.2016	Yes	Yes		
	Nireekshan Engineers and Surveyors	Shandong China Coal	RAD-35 Handheld Radiation Detector	170310122	18.07.2019	07.17.2020	Yes	17.07.2019	Yes	Yes	Technical details are not submitted for the evaluation	The firm is to submit revalidated calibration certificate indicating the technical detail for evaluation and revalidated certificate of membership of Metal Recycling Association of India
		RADCOMM	MSPEC Portale Gamma Spectrometer	200154	14.05.2019	14.05.2020	Yes	03.02.2016	Yes	Yes		

		RADCOMM	TRACEX Explosive Detector - 5 Kits Part No: 510100	16858	NR	NR	NR	15.07.2019	Yes	Yes		
5	APGM Engineering Services Pvt Ltd	SE International	Ranger	R-309592	06.09.2019	05.09.2021	Yes	06.09.2019	Yes	Yes		The firm is to submit revalidated certificate of membership of Metal Recycling Association of India
		Radcom	Mspec	200224	06.09.2019	05.09.2021	Yes	05.01.2019	Yes	Yes		
		Explosives Detection Morphis Technologies	TraceX	16384	NR	NR	NR	30.08.2019	Yes	Yes		
6	Pashupati Consulting and Inspection Agency	Atomtex, Belarus	AT6130	20357	18.06.2014	17.06.2016	Yes	22.09.2014	Yes	Yes	Calibration certificate expired	

		FLIR, Germany	Identifinder2-NG Hand-Held Radionuclide Deive	910383-153	03.05.2019	-	Yes	22.09.2014	Yes	Yes	Calibration certificate not submitted	The firm is to submit revalidated calibration certificate of all instruments clearly indicating its' validity . The firm is also to submit photograph of it's explosive detector and revalidated certificate of membership of Metal Recycling Association of India
		RedXDefence, United States	XCAT Narcotics, Gunshot residue and Explosive Detection System	3455	NR	NR	NR	17.08.2015	Yes	NA		
7	MELT Enterprise LTD	RADCOMM Systems Corp	Mspec Handheld Gamma Ray Spectrometer	200162	08.05.2019	07.05.2020	Yes	08.05.2019	Yes	Yes		The firm is to submit revalidated certificate of membership of Metal Recycling Association of India
		S.E International INC Inspector USB	Portable Radiation Detector	47555	25.07.2019	25.07.2020	Yes	10.07.2016	Yes	Yes		

		S.E International INC Inspector USB	Portable Radiation Detector	47556	25.07.2019	25.07.2020	Yes	10.07.2016	Yes	Yes		
		S.E International INC Inspector USB	Portable Radiation Detector	47557	25.07.2019	25.07.2020	Yes	10.07.2016	Yes	Yes		
		S.E International INC Inspector USB	Portable Radiation Detector	47558	25.07.2019	25.07.2020	Yes	10.07.2016	Yes	Yes		
		TraceX		Lot no 16858 Part No	NR	NR	Yes	19.07.2019	Yes	Yes		
8	Nectar Inspection Services LLC	International Medcom Inc	Radalert 100	73388	02.11.2016	02.11.2020	Yes	15.12.2012	Yes	Yes	Calibration certificate issued for more than 2 yrs validity, Therefore as per the AERB policy, the submitted CC is not acceptable	The firm is to submit revalidated calibration certificate of all its' instruments clearly indicating its' validity and revalidated certificate of membership of Metal Recycling Association of India
		International Medcom Inc	R-100X	X00618	02.11.2016	02.11.2020	Yes	15.07.2014	Yes	Yes		
		S.E International INC	Inspector EXP	28216	29.09.2016	29.09.2020	Yes	28.05.2015	Yes	Yes		
		S.E International INC	Inspector EXP	28214	28.09.2016	28.09.2020	Yes	28.05.2015	Yes	Yes		
		S.E International INC	Inspector EXP	28215	27.09.2016	27.09.2020	Yes	28.05.2015	Yes	Yes		
		S.E International INC	Inspector EXP	28277	28.09.2016	28.09.2020	Yes	10.09.2015	Yes	Yes		
		S.E International INC	Inspector EXP	28359	27.09.2016	27.09.2020	Yes	10.09.2015	Yes	Yes		
		S.E International INC	Inspector EXP	28358	29.09.2016	29.09.2020	Yes	10.09.2015	Yes	Yes		
		RAD COMM	Mspec	200092	18.09.2019	17.09.2020	Yes	15.07.2014	Yes	Yes		
		RAD COMM	Mspec	200085	18.09.2019	17.09.2020	Yes	28.05.2015	Yes	Yes		

		RAD COMM	Mspec	200086	18.09.2019	17.09.2020	Yes	28.05.2015	Yes	Yes		
		RAD COMM	Mspec	200099	18.09.2019	17.09.2020	Yes	28.05.2015	Yes	Yes		
		RAD COMM	Mspec	200103	18.09.2019	17.09.2020	Yes	28.05.2015	Yes	Yes		
		XLT 898 NITON	Thermo Niton Scrap Analysis System	9633	NR	NR	NR	09.06.2012	Yes	Yes		
9	G P & Associates	RADCOMM	Rhandy	100561	05.01.2017	04.01.2019	Yes	05.01.2017	Yes	Yes	Calibration Certificate Expired (+2yrs)+ Annexure- 1 of Application form not filed completely	The firm is to submit revalidated calibration certificate of all instruments and revalidated certificate of membership of Metal Recycling Association of India
		RADCOMM	MSPEC	200267	16.01.2017	16.01.2018	Yes	05.01.2017	Yes	Yes	Calibration Certificate Expired ; Annexure-1 of Application form not filed completely	
			TraceX	510100	NR	NR	NR	14.02.2018	Yes	Yes		

10	CMW Survey & Insepection, The Netherlands	RAE Systems	Gamma Rae	140500057	09.09.2019	08.09.2020	Yes	16.12.2011	Yes	Yes	On IMC Remark: Institution may be asked to provide delivery chalan or any other document for the verification	The S.No. of the instruments are not mentioned on the invoices except in the case of one instrument. The firm is to provide delivery chalan or any other document for the verification of serial no. The firm is also to provide fresh calibration certificate of its instruments clearly indicating the period of vailidity of the calibration and the test results. The firm is to submit revalidated certificate of its membership of
		Gamma Scout	Gamma Rae	63993	19.09.2019	18.09.2020	Yes	23.11.2015	Yes	Yes		

11	Hamilton Steel Logistics Inc.	RADCOMM	RC-2	3335	04.07.2019	-	Yes	15.10.2010	Yes	NA	Calibration certificate neither seal and signed by the agency, therefore CC not acceptable	The firm is to provide calibration certificate duly signed and seal of the calibrating agency and also clearly indicating the vailidity of the calibration and photograph of the instrument .Also revalidated certificate of membership of Metal Recycling Association of India
		RADCOMM	Syclone	50046	29.08.2019	29.08.2020	Yes	10.01.2013	Yes	Yes		
		Spectrex	XD 2 EX-DETECT	18787	NR	NR	NR	01.06.2015	Yes	Yes		
12	Marine Inspection and Logistics International	Exploranium GR-110-E	Exploration radiation detection	3175	08.03.2019		No					Firm is to submit its duly filled ANF 2L form
		Inspector Xtreme USB	S.E International INC	41787	08.03.2019		No	29.05.2015	Yes	Yes		
		Inspector Xtreme USB	S.E International INC	47506	19.11.2018		No			Yes		
		Inspector Xtreme USB	S.E International INC	47507	NA		NA			Yes		
		Inspector Xtreme USB	S.E International INC	41963	08.03.2019		No	08.09.2015	Yes	Yes	No Application submitted	

13	Cayle Aerospace Inc	Thermo Scientific	FH-40G	30223	03.03.2019	02.03.2021	Yes	13.04.2015	Yes	Yes	2 years (NO expiration date)	The firm is to provide calibration certificates of its instruments clearly indicating its vailidity and test results. The firm is also to provide clear photograph of its instrument Sl. No. 50023 and also photograph and invoice of explosive detector. The revailidated membership certificate Metal Recycling Association of India also to be provided
		Thermo Scientific	FH-40G	26183	02.02.2019	01.02.2021	Yes	23.03.2010	Yes	Yes	2 years (NO expiration date)	
		Radcomm	Syclone	50023	Not available			23.03.2010	Yes	S.No. not visible	Test results and date are not mentioned in the submitted certificate	
		Explosive Detector	Fido XT	98447	NR		NR					
14		ZORO Tools	Monitor M4	74975	13.08.2019	13.08.2021	Yes	Serial No not mentioned on the invoice		Yes		The firm is to
	NMCI Inspection & Survey Company Pvt Ltd	Pla Electro	PSDSM 162S	352	01.08.2019	30.09.2021	Yes	Serial No not mentioned on the invoice		Yes		

15	GEO Chem Inspection (Malaysia) SDN. BHD	Rhandy Portable Radiation Detector		100333	Not submitted			17.04.2014	Yes		Calibration certificate not submitted	The firm is to submit calibration certificate of all instruments clearly indicating its vailidity. Invoice & photograph of instruments. Photograph and invoice of the explosive detector and also the certificate of Metal Recycling Association of India
		Rhandy Portable Radiation Detector		100381	Not submitted						Calibration certificate not	
		Medspec Portable Gamma Spectrometer		200091	Not submitted			17.04.2014	Yes		Calibration certificate not	

16	GEO Chem Middle East	RADCOMM	MSpec	200239	14.05.2019		Yes			Yes	Incomplete submission of Annexure-1 of the application form .	The applicant has to submit completed Annexure-1 of the application form. Calibration certificate of its instruments clearly indicating the vailidity period. Also the copy of the invoice all instruments and explosive detector. Also the certificate of its membership of Metal Recycling Association of India
		RADCOMM		200240						Yes		
		NDS 2000		46275						Yes		
		RADCOMM		100430	14.05.2019		Yes			Yes		
		RHANDY		100427	14.05.2019		Yes			Yes		
		SYCLON		50031	31.05.2019		Yes			Yes		
		Morphix Technologies	TraceX	7461						Yes		

17	Meruti Chartered Engineers and Valuers	GAMMA	SCOUT	104243	Date not mentioned on calibration certificate		Yes	13.09.2017	Yes	Yes	Expiration date not mentioned in the CC; and will be treated as expired	The firm is to provide calibration certificate clearly indicating the validity of the certificate The firm is also to provide revalidated membership certificate of Metal Recycling Association of India
		RADCOMM	MSPEC	200343	27.09.2019	26.09.2020	Yes	27.09.2019	Yes	Yes		
		MORPHEX TECHNOLOGIES	TRACE X EXPLOSIVES DETECTION	510100	NR	NR	NR	16.09.2019	Yes	Yes		

18	Nippon Kaiji Kentei Kyokai	Hitachi Aloka Medical, Ltd.	PDR-111	204U9616	05.11.2018	N/A	Yes	24.08.2018	Yes	Yes
		Hitachi Aloka Medical, Ltd.	PDR-111	20238000	02.05.2018	N/A	Yes	24.09.2013	Yes	Yes
		Hitachi Aloka Medical, Ltd.	PDR-111	20238001	02.05.2018	N/A	Yes	24.09.2013	Yes	Yes
		Hitachi Aloka Medical, Ltd.	PDR-111	201K7791	01.08.2018	N/A	Yes	NA	Yes	Yes
		LUDLUM Measurements, Inc.	Model 3 / 44-2	278769	20.06.2018	N/A	Yes	12.04.2011	Yes	Yes
		Hitachi, Ltd.	TCS-1172	205L5683	20.11.2018	N/A	Yes	07.11.2018	Yes	Yes
		Hitachi Aloka Medical, Ltd.	PDR-111	201A4641	11.03.2019	N/A	Yes	NA	Yes	Yes
		Hitachi, Ltd.	TCS-1172	205L5682	20.11.2018	N/A	Yes	07.11.2018	Yes	Yes
		LUDLUM Measurements, Inc.	Model 3 / 44-2	278703	13.06.2018	N/A	Yes	12.04.2011	Yes	Yes
		Hitachi Aloka Medical, Ltd.	PDR-111	20237760	04.04.2018	N/A	Yes	31.10.2013	Yes	Yes

Firm needs to provide the accreditation certificate about the NKKK authorized caribration agency from the comptent authority of their country, for further evaluation clearly indicating the validity of the calibration . The firm is also to provide invoice of the instrument Sl.no.s 201K7791, 51-P22995B, 51-P13055A, P17250479 & 102374 indicating the date of purchase. The firm is also to provide

date not in english

date not in english

NA

		Hitachi Aloka Medical, Ltd.	PDR-111	20238065	30.08.2018	N/A	Yes	20.01.2014	Yes	Yes
		Hitachi, Ltd.	TCS-1172	205W0946	08.03.2018	N/A	Yes	09.02.2018	Yes	Yes
		LUDLUM Measurements, Inc.	Model 3 / 44-2	278692	13.04.2018	N/A	Yes	12.04.2011	Yes	Yes
		LUDLUM Measurements, Inc.	Model 3 / 44-2	278866	12.03.2019	N/A	Yes	12.04.2011	Yes	Yes
		LUDLUM Measurements, Inc.	Model 3 / 44-2	278947	27.06.2018	N/A	Yes	12.04.2011	Yes	Yes
		Hitachi, Ltd.	TCS-1172	205W0947	22.02.2019	N/A	Yes	09.02.2018	Yes	Yes
		Hitachi Aloka Medical, Ltd.	PDR-111	204J1693	16.05.2018	N/A	Yes	30.03.2016	Yes	Yes
		LUDLUM Measurements, Inc.	Model 3 / 44-2	278744	19.05.2018	N/A	Yes	12.04.2011	Yes	Yes
		LUDLUM Measurements, Inc.	Model 3 / 44-2	278841	11.04.2018	N/A	Yes	12.04.2011	Yes	Yes
		Hitachi, Ltd.	TCS-1172	205W0948	27.02.2019	N/A	Yes	09.02.2018	Yes	Yes
		LUDLUM Measurements, Inc.	Model 3 / 44-2	278944	28.03.2018	N/A	Yes	12.04.2011	Yes	Yes
		LUDLUM Measurements, Inc.	Model 3 / 44-2	278852	26.04.2018	N/A	Yes	12.04.2011	Yes	Yes
		Hitachi Aloka Medical, Ltd.	PDR-111	R01209	20.06.2018	N/A	Yes	09.12.2010	Yes	Yes
		Hitachi Aloka Medical, Ltd.	TCS-1172	20160634	08.10.2018	N/A	Yes	12.08.2011	Yes	Yes
		LUDLUM Measurements, Inc.	Model 3 / 44-2	278712	06.10.2018	N/A	Yes	12.04.2011	Yes	Yes
		Hitachi Aloka Medical, Ltd.	TCS-1172	200T3735	29.06.2017	N/A	Yes	12.08.2011	Yes	Yes
		Hitachi Aloka Medical, Ltd.	TCS-1172	202Y7239	29.05.2018	N/A	Yes	03.10.2013	Yes	Yes
		Hitachi, Ltd.	TCS-1172	205W0955	11.03.2019	N/A	Yes	09.02.2018	Yes	Yes
		LUDLUM Measurements, Inc.	Model 3 / 44-2	278871	16.05.2018	N/A	Yes	12.04.2011	Yes	Yes
		LUDLUM Measurements, Inc.	Model 3 / 44-2	278928	16.05.2018	N/A	Yes	12.04.2011	Yes	Yes
		Hitachi, Ltd.	TCS-1172	205L0376	03.07.2018	N/A	Yes	28.06.2017	Yes	Yes



19	Covenant Insurance Surveyors & Loss Assessors Pvt Ltd	NA	NA	NA	NA	N/A	NA	NA	NA	NA	Details of equipment are not mentioned in the Annexure-1 form of the application.	The applicant has to submit completed Annexure-1 of the application form. Calibration certificate of its instruments clearly indicating the vailidity period. Also the copy of the invoice all instruments and explosive detector. Also the certificate of its membership of Metal Recycling Association of India
20	Premier Exim	Rhandy	Radcomm	100654	10.09.2019	09.09.2020	Yes	10.09.2019	Yes	Yes	Incomplete submission of Annexure-1 of the application form .	The firm is to provide completed application along with Annexure-1. The firm is also to provide membership certificate of Metal Recycling Association of India
		MSPEC	Spectrometer	200336	10.09.2019	09.09.2020	Yes	10.09.2019	Yes	Yes		

		Explosive	TraceX		N/R	N/R	N/R	09.09.2019	Yes	Yes		
21	Ruswal Global LLC	Rostechnika LLC	ecolog mini	837	13.06.2019	13.06.2020	Yes	NA	NA	Yes	Spectrometer details are not mentioned in the application form. Calibration certificate are in Russian Language and non certified translated copy (English version) submitted. *** On IMC Remark: Institution may be asked to provide delivery chalan or The firm is to submit duly completed ANF-2L form along with Annexure-1. Copy of the invoices clearly indicating the serial no. of the instruments or provide delivery chalan or other document for the verification. Calibration certificate of the instruments clearly indicating the test result and its vailidity. The calibration certificates of the instruments should not be more than 2 years old. All translated documents should be self attested and the invoice and photograph of explosive detector is also to be provided. The firm is to submit its revailadated certificate of its membership of Metal Recycling Association of India	
		Rostechnika LLC	ecolog	840	13.06.2019	13.06.2020	Yes	NA	NA	Yes		
		Rostechnika LLC	ecolog	847	13.06.2019	13.06.2020	Yes	NA	NA	Yes		

		Rostekhnika LLC	ecolog for a	518	13.06.2019	13.06.2020	Yes	NA	NA	Yes		
--	--	-----------------	--------------	-----	------------	------------	-----	----	----	-----	--	--

FD

		Rostekhnika LLC	ecolog mini	933	NA		NA	NA	NA	Yes	Test results are not mentioned in the ...
--	--	-----------------	-------------	-----	----	--	----	----	----	-----	---

12

		Rostechnika LLC	ecolog mini	937	NA		NA	NA	NA	Yes	calibration certificate.	

12

22	NMCI Inspection & Survey (MTIUS) LTD	RADCO MM SYSTEM S	R HANDY	100544	08.08.201 6		Yes	09.08.201 6	Yes	Yes	Calibration certificate is not submitted	The applicant is to submit calibration certificates of its instruments clearly indicating their validity. The Uncertainty level (error in %) of Instruments should not be more than the acceptance limit. The firm is also submit copy of the invoices clearly indicating the serial no. of the instruments or provide delivery chalan or other document for the verification of serial no. The details of explosive detector, photograph and invoice is also to be provided. The revalidated membership certificate of International Federation of Inspection Agencies
		RADCO MM SYSTEM S	R HANDY	100545	28.08.201 6		Yes	09.08.201 6	Yes	Yes	Calibration certificate is not submitted	
		GS Geuger Group	Gamma Scout	103646	13.02.201 8	12.0 3.20 20	Yes	08.05.201 7	NA	Yes	On IMC Remark: Institution may be	SI No. not on invoice

GS Geuger Group	Gamma Scout	103649	03.03.2018	02.03.2020	Yes	08.05.2017	NA	Yes	asked to provide delivery chalan or other document for the verification	Sl No. not on invoice
GS Geuger Group	Gamma Scout	104402	03.03.2018	02.03.2020	Yes	08.05.2018	NA	Yes		Sl No. not on invoice
RADCOMM SYSTEMS	R HANDY	100609	04.05.2019		Yes	13.02.2018	Yes	Yes	Test results are not mentioned in the calibration certificate.	
RADCOMM SYSTEMS	R HANDY	100610	04.05.2019		Yes	13.02.2018	Yes	Yes		
GS Geuger Group	Gamma Scout	67808			Yes	09.01.2018	NA	Yes	Calibration certificate is not submitted	Sl No. not on invoice
GS Geuger Group	Gamma Scout	104696	05.09.2018	04.09.2020	Yes	31.05.2018	NA	Yes	On IMC Remark: Institution may be asked to provide delivery chalan or other document for the verification	Sl No. not on invoice
GS Geuger Group	Gamma Scout	104698	05.09.2018	04.09.2020	Yes	31.05.2018	NA	Yes		Sl No. not on invoice
PLA	PSDM 162S	347	14.05.2019	13.05.2021	Yes	15.05.2019	Yes	Yes		
PLA	PSDM 162S	348	14.05.2019		Yes	15.05.2019	Yes	Yes	Calibration certificate is not submitted	

NA

		RADCOMM	MSPEC	200222	04.05.2019	03.05.2020	Yes	09.08.2016	Yes	Yes	Uncertainty level (error in %) of Instruments is more than the acceptance limit.	
		RADCOMM	MSPEC	200223	04.05.2019	03.05.2020	Yes	09.08.2016	Yes	Yes	Uncertainty level (error in %) of Instruments is more than the acceptance limit.	
	Addition of instruments and Area of Operation											
23	Baltic Control International Ltd UK	Polon Ekolab	EKO-P	698/ 2018	08.05.2019	N/A	Yes	06.04.2018	Yes	Yes	error percentage is not mentioned in the certificate, however as per the result the error percentage is more prescribed limit.	Firm is to submit calibration certificate indicating error percentage, which is not mentioned in the certificate, however as per the result the error percentage is more prescribed limit. The validity of the Calibration should be clearly indicated on the calibration certificate.

		Polon Ekolab	EKO-P	699/ 2018	08.05.2019	N/A	Yes	06.04.2018	Yes	Yes	error percentage is not mentioned in the certificate, however as per the result the error percentage is more prescribed limit.	
		Polon Ekolab	EKO-P	701/ 2018	08.05.2019	N/A	Yes	06.04.2018	Yes	Yes	error percentage is not mentioned in the certificate, however as per the result the error percentage is more prescribed limit.	
		Polon Ekolab	EKO-P	702/ 2018	08.05.2019	N/A	Yes	06.04.2018	Yes	Yes	error percentage is not mentioned in the certificate, however as per the result the error percentage is more prescribed limit.	
		Polon Ekolab	EKO-P	820/ 2017	01.03.2019	N/A	Yes	31.05.2017	Yes	Yes	error percentage is not mentioned in the certificate, however as per the result the error percentage is more prescribed limit.	



		Polon Ekolab	EKO-P	821/ 2017	01.03.2019	N/A	Yes	31.05.2017	Yes	Yes	error percentage is not mentioned in the certificate, however as per the result the error percentage is more prescribed limit.	
		Polon Ekolab	EKO-P	822/ 2017	01.03.2019	N/A	Yes	31.05.2017	Yes	Yes	error percentage is not mentioned in the certificate, however as per the result the error percentage is more prescribed limit.	
		Polon Ekolab	EKO-P	823/ 2017	01.03.2019	N/A	Yes	31.05.2017	Yes	Yes	error percentage is not mentioned in the certificate, however as per the result the error percentage is more prescribed limit.	
		RADCOMM	Mspec	200230	05.03.2019	04.03.2021	Yes	30.09.2016	Yes	Yes		
		Ecolog Mini	OOO Rostekhnika	732	21.02.2019	N/A	Yes	03.01.2017	Yes	Yes		

		Hitachi, Ltd.	TCS-1172	205L0375	03.07.2018	N/A	Yes	28.06.2017	Yes	Yes
		LUDLUM Measurements, Inc.	Model 3 / 44-2	278719	10.04.2018	N/A	Yes	12.04.2011	Yes	Yes
		Hitachi, Ltd.	TCS-1172	205L0374	06.04.2018	N/A	Yes	28.06.2017	Yes	Yes
		LUDLUM Measurements, Inc.	Model 3 / 44-2	278817	06.04.2018	N/A	Yes	12.04.2011	Yes	Yes
		LUDLUM Measurements, Inc.	Model 3 / 44-2	278814	31.05.2018	N/A	Yes	12.04.2011	Yes	Yes
		Hitachi, Ltd.	TCS-1172	205L0373	03.07.2017	N/A	Yes	28.06.2017	Yes	Yes
		LUDLUM Measurements, Inc.	Model 3 / 44-2	278777	31.05.2018	N/A	Yes	12.04.2011	Yes	Yes
		LUDLUM Measurements, Inc.	Model 3 / 44-2	278833	02.04.2018	N/A	Yes	12.04.2011	Yes	Yes
		LUDLUM Measurements, Inc.	Model 3 / 44-2	278752	14.08.2018	N/A	Yes	12.04.2011	Yes	Yes
		Hitachi, Ltd.	TCS-1172	205L5746	05.02.2019	N/A	Yes	28.06.2017	Yes	Yes

Self certified calibration certificate of equipments from NKKK are submitted. Applicant needs to provide the accreditation certificate about the NKKK authorized caribration agency from the comptent authority of their country, for further evaluation.

		Inspector Xtreme USB	S.E International INC	41732	08.03.2019		No	08.09.2015	Yes			
		Inspector Xtreme USB	S.E International INC	47118	08.03.2019		No	09.03.2016	Yes	Yes		
		Inspector Xtreme USB	S.E International INC	6510	29.01.2019		Yes					
		Inspector Xtreme USB	S.E International INC	20817	15.01.2019		Yes			Yes		
		Inspector Xtreme USB	S.E International INC	41790	14.03.2019		Yes	29.05.2015	Yes	Yes		
		Inspector Xtreme USB	S.E International INC	41793	20.02.2019		Yes	29.05.2015	Yes	Yes		
		Inspector Xtreme USB	S.E International INC	48387	13.02.2019		Yes			Yes		
		Radiation Alert Ranger SE International	S.E International INC	6037	08.03.2019					Yes		
		Radiation Alert Ranger SE International	S.E International INC	6038	08.03.2019					Yes		
		Radiation Alert Ranger SE International	S.E International INC	R307296	05.10.2018		Yes			Yes		
		Radiation Alert Ranger SE International	S.E International	R309071	22.03.2019		Yes			Yes		
		Radiation Alert Ranger SE International	S.E International	R309072	21.03.2019		Yes			Yes		
		FLIR radiation detection	Identifier 2	910383-36	NA		NA	04.03.2013	Yes	Yes		
		FLIR radiation detection	Identifier 2	4628-61	NA		NA	04.03.2013	Yes	Yes		

NA

		Radex	Quarta	RD1706	19.04.2019	18.04.2021	Yes	08.04.2015	Yes	Yes	2 years (NO expiration date)	
		Gamma Scout	Gamma Scout	64059	23.04.2019	22.04.2021	Yes	08.04.2015	Yes	Yes	2 years (NO expiration date)	
		Gamma Scout	Gamma Scout	64056	24.04.2019	23.04.2021	Yes	08.04.2015	Yes	Yes	2 years (NO expiration date)	
		Gamma Scout	Gamma Scout	103004	06.02.2019	05.02.2021	Yes	29.05.2015	Yes	Yes	2 years (NO expiration date)	

		Gamma Scout	Gamma Scout	64375	23.07.2019		Yes	12.05.2015	Yes	Yes	Test results are not mentioned in the calibration certificate	
		Gamma Scout	Gamma Scout	103080	17.09.2019	16.09.2021	Yes	13.07.2015	Yes	Yes	2 years (NO expiration date)	
		Gamma Scout	Gamma Scout	105257	27.12.2018	26.12.2020	Yes	14.12.2018	Yes	Yes	2 years (NO expiration date)	
		Gamma Scout	Gamma Scout	105254	27.12.2018	26.12.2020	Yes	14.12.2018	Yes	Yes	2 years (NO expiration date)	
		RAE Systems	QRAE II Diffusion	155-113765	09.09.2019	08.09.2020	Yes	07.10.2016	Yes	Yes	2 years (NO expiration date)	
		RAE Systems	PGM 7300	590-000121	09.09.2019	08.09.2020	Yes	02.12.2018	Yes	Yes	2 years (NO expiration date)	
		Radex	Quarta	05172AA8551476	26.04.2019	25.04.2020	Yes	08.02.2018	Yes	Yes	2 years (NO expiration date)	
		Radex	Quarta	05172A8551463	20.02.2019	19.02.2020	Yes	08.02.2018	Yes	Yes	2 years (NO expiration date)	
		Radex	Quarta	02182AA8551742	27.12.2018	26.12.2020	Yes	14.12.2018	Yes	Yes	2 years (NO expiration date)	
		Gamma Scout	Gamma Scout	74148	19.09.2019	18.09.2021	Yes	09.09.2019	Yes	Yes	2 years (NO expiration date)	
		Gamma Scout	Gamma Scout	74158	19.09.2019	18.09.2021	Yes	09.09.2019	Yes	Yes	2 years (NO expiration date)	
		Target	IronMAN	1120-3536	27.02.2019	26.02.2021	Yes	19.09.2018	Yes	Yes	2 years (NO expiration date)	S.No of the instrument mentioned on invoice
		Exploranium	GR130	1414+GB-B	16.07.2019	15.07.2021	Yes	07.02.2016	Yes	Yes	2 years (NO	

		Pla Electro	PSDSM 162S	353	01.08.2019	30.09.2021	Yes		Invoice Not Available	Yes	On IMC Remark: Institution may be asked to provide delivery chalan or other document for the verification	submit invoice clearly indicating the serial no. of the instruments or provide delivery challan or other document for the verification of instruments. The firm is to submit revalidated calibration certificate
		RADCOMM	MSPEC	200090	04.05.2019	03.05.2021	Yes	04.04.2014	Yes	Yes		
		IndySurplus	Sabre 2000	20285	NR	NR	Yes	04.08.2014	Yes	Yes		



		LUDLUM Measurements, Inc.	Model 3 / 44-2	278739	28.06.2017	N/A	Yes	12.04.2011	Yes	Yes
		Hitachi Aloka Medical, Ltd.	TCS-1172	204N9448	16.05.2018	N/A	Yes	20.07.2016	Yes	Yes
		LUDLUM Measurements, Inc.	Model 3 / 44-2	324175	07.05.2018	N/A	Yes	NA	Yes	Yes
		Hitachi Aloka Medical, Ltd.	PDR-111	20487900	20.08.2018	N/A	Yes	18.09.2015	Yes	Yes
		Hitachi Aloka Medical, Ltd.	PDR-111	20487904	30.08.2018	N/A	Yes	25.09.2015	Yes	Yes
		LUDLUM Measurements, Inc.	Model 3 / 44-2	278773	20.02.2019	N/A	Yes	12.04.2011	Yes	Yes
		SEIKO EG&G CO., LTD.	GEM25-70	51-P22995B	23.05.2017	N/A	NA	NA	NA	NA
		SEIKO EG&G CO., LTD.	GEM30-70	51-P13055A	23.05.2017	N/A	NA	NA	NA	NA
		Rigaku Analytical Devices, Inc.	Progeny ResQ 1064	P17250479	02.09.2019	N/A	NA	NA	NA	NA
		Thermo Fisher Scientific Inc.	XL-3t-9805	102374	02.09.2019	N/A	NA	NA	NA	NA

Not in english
Photograph and other document of explosive detector not attached
Photograph and other document of explosive detector not attached